

# The Megger guide to insulating oil dielectric breakdown testing



**Megger** <sup>®</sup>

[WWW.MEGGER.COM](http://WWW.MEGGER.COM)





## Contents

1. Introduction .....	3
2. What are the methods for testing oil? .....	3
3. What is oil dielectric breakdown voltage testing? .....	4
4. Who needs to perform dielectric breakdown voltage tests? .....	4
5. Which types of insulating oil can be tested? .....	4
6. Why, when and how often is oil testing performed? .....	6
7. What are the test standards and their differences? .....	6
7.1 ASTM standards (USA) .....	8
7.2 IEC standards (international) .....	9
8. How should an oil sample be taken? .....	10
9. Preparing the test vessel .....	12
10. Why measure oil sample temperature? .....	13
11. How do I know whether my test results are valid? .....	13
12. How do I know if my fluid has passed the test? .....	14
12.1 USA – ASTM and IEEE standards .....	14
12.2 International IEC standards .....	15
13. Can new oil fail a dielectric breakdown test? .....	16
14. Can I verify my test instrument performance? .....	17
15. How do oil test sets detect dielectric breakdown? .....	17
15.1 Testing silicone oil .....	18
16. Choosing an oil dielectric breakdown voltage test set .....	18
16.1 On-site versus laboratory testing .....	18
16.2 General instrument selection considerations .....	18
16.3 Selection considerations specific to laboratory instruments .....	24
16.4 Selection considerations specific to portable instruments .....	25
17. Megger OTS range summary .....	27

## Acknowledgments

Megger gratefully acknowledges the support of John Noakes of TJ/H2b Analytical Services for his help in compiling this publication.

## 1. Introduction

Oils that combine a high flashpoint with high dielectric strength have long been used as an insulating medium in transformers, switchgear and other electrical apparatus. To ensure that the dielectric strength of the oil does not deteriorate however, proper maintenance is essential, and the basis of proper maintenance is testing.

For over 100 years, Megger has been a world leader in the development and manufacture of test equipment for electrical power applications. The famous Megger trademark was first registered in 1903, and jealously guarded by the company. Megger's experience in the design and production of oil test sets also reaches back to the early 20th century, when pioneering equipment was produced by Foster Transformers, a company that became part of Megger Group in 1968.

Today, all of Megger's oil test set design, development and manufacture is carried out at the company's manufacturing facility in Dover, England.

## 2. What are the methods testing oil?

For in-service equipment in particular, there are many test techniques for evaluating the condition of the insulating oil. If the technique of dissolved gas analysis is excluded, oil tests can be divided into two basic groups.

The first group includes tests that are concerned with the immediate condition and acceptability of the insulation in an item of electrical equipment. This group includes dielectric breakdown voltage testing as well as moisture measurement by the Karl Fischer (KF) method, and determination of insulation condition by measuring the dielectric dissipation factor.

### The Megger OTS range of dielectric breakdown voltage test sets



### The Megger KF range of moisture content test sets



The second group includes tests that look at the degree of degradation and aging of the equipment's insulation system. These tests include interfacial tension, acidity (neutralisation value), resistivity and visual determination of colour and appearance of the insulating oil.

### Typical oil colour specimens


The darker the colour the more contaminants present

Good ..... Fair ..... Bad



While all of these tests are useful in particular circumstances, the most convenient and most universally applicable is dielectric breakdown voltage testing, which is why it has been selected as the subject of this technical guide.

## 3. What is oil dielectric breakdown voltage testing?



Put simply, a dielectric breakdown voltage test is a measure of the electrical stress that an insulating oil can withstand without breakdown. The test is performed using a test vessel that has two electrodes mounted in it, with a gap between them. A sample of the oil to be tested is put into the vessel and an ac voltage is applied to the electrodes. This voltage is increased until the oil breaks down – that is, until a spark passes between the electrodes. The test voltage is then immediately turned off. The voltage at which breakdown occurred is the test result, and is typically evaluated by comparing it with guidelines set out in various standards, or in the oil manufacturer's specifications.

The exact method of performing the test is determined by the standard that is being used, as will be explained later. The standard typically defines parameters such as the size and shape of the electrodes, the gap between them, the rate at which the test voltage is increased, how many times the test is repeated and whether or not the oil is stirred during the test.

## 4. Who needs to perform dielectric breakdown voltage tests?

There are many types of organisation that benefit from carrying out tests on transformer oil. These include:

- Utility contractors (principally in substations)
- Utility companies (principally in power stations)
- Rail companies (locomotive HV step-down transformers and switchgear)
- Oil test laboratories (providing testing services)
- Transformer and switchgear manufacturers (quality control of oil)
- Oil companies (testing new oil during manufacture)

## 5. Which types of insulating oil can be tested?

While the generic term 'oil' is almost universally used to describe insulating fluids, there are currently five different types of insulating fluid in common use. These are:

- Mineral oil
- High molecular weight hydrocarbon (HMWH) fluids
- Silicone fluids
- Synthetic ester fluids
- Natural ester (vegetable oil) fluids

All of these oil types can be tested for dielectric breakdown voltage and tested with Megger OTS range test sets.

Mineral oil is the most common insulating fluid and has been in use since the late 19th century. There are many mineral oil filled transformers that have been in continuous use for more than 50 years. Mineral oils are refined from either naphthenic crude or more recently, from paraffinic crude.

HWMH, silicon, synthetic ester and natural ester fluids are more recent developments and are often preferred because they are much less flammable than mineral oil. ASTM D5222 specifies that for insulating fluids to qualify as 'less flammable' they must have a fire point of at least 300 °C.

The five fluids differ significantly in the way they behave in the presence of moisture. Mineral oil is the least satisfactory, and even small amounts of water significantly reduce its breakdown voltage. Silicone fluid is also quickly affected by small amounts of moisture, whereas ester fluids behave very well in the presence of moisture and can typically maintain a breakdown voltage of greater than 30 kV with more than 400 ppm water content. This is one of the reasons that esters last much longer in service.

This table is not exhaustive, but does give a good indication of the types of insulating fluid that are used in various applications. It also shows which types of fluid can be tested with the Megger OTS range of dielectric breakdown voltage test sets.

Equipment	Fluid Type		Example/Sub-type	Can be tested with Megger OTS range?	
Capacitors	Synthetic aromatic hydrocarbons		PXE	Yes	
	Aromatic esters		Various types	Yes	
Medium and high voltage cables	New	Synthetic hydrocarbons	Polybutenes	Yes	
	Old	Mineral oil	Various types	Yes	
Bushings	Mineral oil		Various types	Yes	
Oil filled circuit breakers	Mineral oil		Various types	Yes	
Transformers	Mineral oil		Shell Diala AX	Yes	
	Perfluorocarbon (PFC)		3M PF-5060	Yes	
	Low flammability fluids	High molecular weight (HMW) oil		Various types	Yes
		Silicone		Dow Corning 561	Yes
		Synthetic hydrocarbons		Polyalohaolefins (PAOs)	Yes
		Synthetic polyol esters		Envirotemp® 200	Yes
		Vegetable oils - natural ester		Envirotemp® FR3	Yes
		Hydrofluorocarbon		Vertrel® VX	Yes
	Old fluids	PCBs - Polychlorinated biphenyls		Askarel® Pyranol® Phenochlor®	No - Harzardous - requires special handling
		Tetrachloroethylene/perchloroethylene (PCE)		Askarel® (contained 50%) Wecosol®	No - Harzardous - requires special handling
		Gases	Sulphur Hexafluoride	SF6	No
Old gases	Freon R-113		Vapotrans	No	
LTC (Load Tap Changers)	Mineral oil		Various types	Yes	

## 6. Why, when and how often is oil testing performed?

### 6.1 Why and when?

The dielectric breakdown voltage test is a relatively quick and easy way of determining the amount of contamination in insulating oil. Usually the contaminant is water, but it can also be conductive particles, dirt, debris, insulating particles and the by-products of oxidation and aging of the oil.


For in-service equipment, the dielectric breakdown voltage test offers a useful and convenient way to detect moisture and other contamination in the oil before it leads to a catastrophic failure. The information gained from the test can also be used as an aid to:

- Predicting the remaining life of a transformer
- Enhancing operational safety
- Preventing equipment fires
- Maintaining reliability

Dielectric breakdown voltage testing is also carried out on new oil before it is used to fill equipment, and as part of the acceptance testing for deliveries of new and reprocessed oil.

### 6.2 How often?

Dielectric breakdown voltage testing is an important element in the maintenance programme of any item of oil-insulated electrical equipment. However, to get the maximum benefit from this type of testing, Megger strongly recommends that the oil is tested at least once a year and preferably twice a year. The results should be recorded, as trending the data will make it easier to identify sudden or unexpected changes. If a sudden change in the results is found, the transformer can be inspected for leaks, the oil level can be checked and the water content of the oil evaluated. If contamination is confirmed, it will often be possible to dry and filter the oil, thereby reconditioning it rather than having to replace it with expensive new oil.



## 7. What are the test standards and their differences?

There are many test standards for insulating liquids, but they are derivatives of three main standards. Two of these are from ASTM International (USA) and the other is from the IEC (Europe). These main standards are:

- ASTM D877 – Standard Test Method for Dielectric Breakdown Voltage of Insulating Liquids Using Disk Electrodes.
- ASTM D1816 – Standard Test Method for Dielectric Breakdown Voltage of Insulating Oils of Petroleum Origin Using VDE Electrodes.
- IEC 60156 Insulating Liquids – Determination of the breakdown voltage at power frequencies – Test method.
- There is also a Japanese standard not based on these ASTM or IEC standards.

The following table is not exhaustive, but it shows the main differences between the ASTM and IEC standards.



Standards		ASTM D1816	ASTM D 877		IEC 60156
			Procedure A	Procedure B	
Origin		USA	USA	USA	Europe
Electrodes	Shape				
	Gap size	2 mm or 1 mm*	2.54 mm	2.54 mm	2.5 mm
Oil sample stirring	Impeller	yes	not stirred	not stirred	optional
	Magnetic bead	no option			optional
Laboratory test temperature	Liquid	At ambient - must record	20 - 30 °C must record temperature as collected and when tested	20 - 30 °C must record temperature as collected and when tested	15 - 25 °C for referee tests
	Ambient	20 - 30 °C	Must record	Must record	Within 5 °C of oil sample
Outside test temperature	Liquid	At ambient - must record	Must record	Must record	15 - 25 °C
	Ambient	Referee tests 20 - 30 °C	Must record	Must record	Within 5 °C of oil sample
Test voltage	Rise rate	0.5 kV/s	3 kV/s	3 kV/s	2 kV/s
	Frequency	45 - 65	45 - 65	45 - 65	45 - 62
Breakdowns	Definition	<100 V	<100 V	<100 V	4 mA for 5 ms
	Number in sequence	5**	5*	1 - 5 different samples	6
	Time between breakdown	1 to 1.5 min	1 min	n/a	2 min
Test voltage switch off time following breakdown	Normal (e.g. mineral oil)	Not specified	Not specified	Not specified	<10 ms
	Silicon oil	Not specified	Not specified	Not specified	<1 ms
Time between filling and start of test		3 - 5 min	2 - 3 min	2 - 3 min	2 min
Equivalent standards (adopted into)		None	None	None	BS EN 60156 CEI EN 60156 IRAM 2341 UNE EN 60156 FN EN 6056 SABE EN 60156 VDE0370 part 5 AS1767.2.1 PA SEV EN 60156 NRS 079-1* IS6729*
Notes on testing silicon oil		Can be used provided discharge energy in sample <20 mj		Can be used if modified in accordance with D2225 if procedure A cannot be used	OK if test instrument can comply with voltage switch off time requirements
Special conditions		* If breakdown does not occur at 2 mm, reduce gap to 1 mm ** Tests must be repeated if range of BD voltages recorded are more than 120% of mean with 1 mm electrode gap and 92% of mean with 2 mm electrode gap		*Tests must be repeated if range of BD voltages recorded are more than 92% of mean. If range of 10 BD voltages is more than 151% investigate why	Expected range of standard deviation/ mean ratio as a function of the mean provided as a chart
Comments		Test vessel requires cover or baffle to prevent air from contacting circulating oil	Used if any insoluble breakdown products in oil completely settle between breakdown tests	Used if any insoluble breakdown products do not settle between breakdown tests	*With some stand/stir timing differences. Test cell/vessel must be transparent. Reconditioned/reclaimed oil to BS148 is tested to IEC60156 following update in 2009.

There are many standards based on IEC 60156 and in addition Japanese standard JIS C2101 also includes requirements for dielectric breakdown voltage testing. JIS C2101 calls for spherical electrodes similar to those specified by IEC 60156, but requires a different sequence of five breakdowns. JIS C2101-99 (M) for mineral oils calls for two oil samples to be tested each with a sequence of five breakdown tests. JIS C2101-99 (S) for silicon oils calls for a sequence five tests but each test is performed on a different sample of oil.

## 7.1 ASTM standards (USA)


There are two standard test methods from ASTM International. The first is ASTM D877-02 (latest edition 2002) – Standard Test Method for Dielectric Breakdown Voltage of Insulating Liquids Using Disk Electrodes. The second is ASTM D1816-04 (latest edition 2004) – Standard Test Method for Dielectric Breakdown Voltage of Insulating Oils of Petroleum Origin Using VDE Electrodes. Although this is essentially an American standard, it borrows from VDE, a German standards organisation.

### ASTM D877

ASTM D877 is an older standard, and is generally not very sensitive to the presence of moisture. For that reason it is not widely used for in-service applications. In 2002 the IEEE revised C51.106, Guide for the Acceptance and Maintenance of Insulating Oil in Equipment the values from D877 were removed from their criteria for evaluating in-service oil in transformers. Generally ASTM 877 is recommended only for acceptance testing of new oil received from a supplier, either in bulk loads or containers, to ensure the oil was correctly stored and transported. Typically a minimum breakdown value of 30 kV is specified.

The ASTM D877 standard specifies the use of disc shaped electrodes that are 25.4 mm (1 inch) is diameter and at least 3.18 mm (0.125 inch) thick. These electrodes are made of polished brass and are mounted to have their faces parallel and horizontally in line in the test vessel. The edges are specified to be sharp with no more than a 0.254 mm (0.010 inch) radius. The sharp edges should be regularly inspected to ensure that they have not become too rounded. Excessively rounded edges will have the effect of falsely raising the breakdown voltage, possibly passing oil that should have failed the test. It is also important that the electrodes are kept very clean, with no pitting or signs of corrosion, otherwise breakdown values can be falsely low. Instrument users should regularly inspect electrodes, cleaning and polishing as required.

### The Megger OTS instrument test set up screens for D877:




### ASTM D1816

ASTM D1816 has become widely used over the years, even being used outside the standard's stated application of petroleum origin insulating oils and viscosity limits. D1816 is more sensitive than D877 to moisture, oil aging and oxidation, and is more affected by the presence of particles in the oil. When the IEEE revised C51.106 in 2002, breakdown voltage limits for new and in-service oil using D1816 were added.

ASTM D1816 specifies the use of mushroom-shaped electrodes 36 mm in diameter. As with D877, the electrodes are made of brass must be polished to be free of any etching, scratching, pitting, or carbon accumulation. The oil is stirred throughout the test sequence, and a two-bladed motor-driven impeller is specified. The standard prescribes the impeller dimensions and pitch as well as the operating speed, which must be between 200 rpm and 300 rpm. To prevent air coming into contact with the circulating oil, the test vessel must have a cover or baffle.

The D1816 standard, although generally accepted as more useful than D877, has got one significant limitation: when testing in-service oil this test method is very sensitive to dissolved gases. Excessive amounts of gas in the oil can lower the test results to the point that a perfectly good sample of oil, with low moisture and particle content, will fail the test. It is important to bear this in mind when testing oil from small gas blanketed transformers and even, in some cases, from free-breathing transformers.

**The Megger OTS instrument test set up screen for D1816:**



The insulating fluid manufacturer will usually quote typical breakdown values for both new and in-service fluid in the insulating fluid data sheet. In addition the test standards will refer to another oil condition standard that will provide guidance as to what is acceptable. Test houses will also provide guidance as to what is acceptable.

**7.2 IEC standards (International)**

The International Electrotechnical Commission (IEC) defines only one method for oil breakdown dielectric voltage testing. This is IEC 60156 Insulating Liquids – Determination of the Breakdown Voltage at Power Frequencies – Test Method.


**IEC 60156**

IEC 60156 is an international standard that appears in many forms as IEC member national committees from various countries have adopted it. Examples are British Standard BS EN 60156 and German VDE 0370 part 5. IEC 60156 specifies the use of (either spherical, or) mushroom (shaped) electrodes the same those used in the ASTM D1816 standard. The IEC standard differs in a number of ways from D1816, but the main difference is the IEC standard allows the optional use of a stirring impeller, the use of a magnetic bead stirrer or even no stirring at all. The standard states that differences between tests with or without stirring have not been found to be statistically significant. The use of a magnetic stirrer is only permitted when there is no risk of removing magnetic particles from the oil sample under test. When oil is used as a coolant and therefore circulating it would be stirred during testing. For example oil from a transformer normally circulates if it is used as coolant, so an oil sample taken here would normally be stirred to ensure the best chance of detecting particle contamination. Oil from a circuit breaker is normally static in use so particles would naturally fall to the bottom where they are unlikely to cause a problem. So in static use applications, an oil sample would usually not be stirred.

The dielectric breakdown values from the IEC 60156 method are usually higher than those from the ASTM methods. Possibly this is in part because of the differences in voltage ramp up speed and electrode gap compared with D1816, and electrode shape compared with D877. (The IEC electrode shape provides a more uniform electric field). The result is that for well-maintained transformers the breakdown voltages may be higher than a 60 kV test instrument can reach.

This may not be a problem when evaluating new oil from a supplier or even for in-service oil, but often an actual breakdown voltage value is required. When testing to IEC 60156, therefore, the use of an instrument capable of applying a higher voltage is advisable. As with D1816, dissolved gas in the oil sample may reduce breakdown values but the effect is much less pronounced than with the IEC 60156 standard.

Some Megger OTS instrument test set up screen examples for IEC60156:



## 8. How should an oil sample be taken?

Two things are particularly important when taking oil samples. The first is to ensure that the proper sampling procedure is followed, and the second is to ensure that all of the essential information is properly recorded.

If the sample is to be sent to a test house for testing, the test house should be able to advise on the information needed, but it's important to bear in mind that the condition diagnosis will only be as good as the information supplied. The test house should also advise on the volume of the sample, and the type of container to use.

For oil samples from transformers, this information that oil test laboratories generally require is:

- Description of the sample
- List of tests to be performed
- Transformer name plate information
- Type of transformer
- Type of insulating fluid
- Any leaks noted
- Insulating fluid service history (has it been dried, etc)
- Transformer service history (has it been rewound, etc)
- Type of breather
- Type of insulation, including temperature rise rating
- Details of cooling equipment (fans, radiators, etc)
- Temperature of top of fluid, read from gauge
- Actual fluid temperature measured
- Fluid level
- Vacuum and pressure gauge readings

For load tap changers, it is also advisable to record the counter reading, the selector range and the sweep range.

Sampling should be performed in accordance with the appropriate standard, and is not, therefore, discussed in detail in this technical guide.

In the USA, there are two standards for sampling:

- D923 – Standard Practices for Sampling Electrical Insulating Liquids
- D3613 – Standard Practice for Sampling Electrical Insulating Oils for Gas Analysis and Determination of Water Content

Internationally, there are two further sampling standards:

- IEC 60475 Ed. 2.0 – Method of Sampling Insulating Liquids

- IEC 60567 Ed. 3.0 – Oil-filled electrical equipment – Sampling of gases and of oil for analysis of free and dissolved gases – Guidance

The IEC standards should be consulted together, especially as part of IEC 60567 has been transferred to IEC 60475.

## 8.1 Hints and tips for taking oil samples

For a sample to be truly useful, it must be representative of the oil in the equipment. This means that cleanliness is extremely important.

- Samples are normally drawn from a drain valve or sampling cock. This must be cleaned both inside and out before the sample is taken to ensure that dirt does not fall into the sampling container.
- The drain valve is at the bottom of the equipment, where all of the sludge, water and contaminant particles collect. It is important therefore, to flush the system thoroughly to ensure that the sample is drawn from the main bulk of the oil. This may involve removing two litres of oil, and even more if the equipment has been out of service for some time.
- **Do not** be tempted to use old engine oil bottles, even for a few p.p.m. of engine oil will cause the sample to fail a breakdown test.
- **Do** let the oil flow down the side of the sample bottle, or use a clean tube run to the bottom of the bottle; it will prevent air being mixed with the oil.
- **Do** store the oil samples in glass or clear plastic bottles in the dark, mineral oil will deteriorate if exposed to UV light.

## Safety

- Before taking samples, ensure that you have all of the required permissions and permits
- Have everything you need to lock out/tag out to hand
- Make sure that the PCB (polychlorinated biphenyl) content of the oil, if any, is known and that the equipment is labelled. PCB is very hazardous and requires special handling
- Use all of the correct personal protective equipment (PPE) and correctly rated tools
- Check the area for electrical and tripping hazards
- Check for wildlife – snakes, bees, etc like transformers!
- Check that the transformer is under positive pressure – are the pressure gauges reliable? Could they be blocked or broken? NEVER try to take a sample from a transformer under negative pressure. Air could be drawn into the transformer and cause it to fail.

## Sampling equipment

- Take extra sample bottles and syringes – they're often needed
- Ensure that the sample bottle seals are airtight
- Use only ground glass syringes
- If rubber hose is used, discard after each sample is taken

### A sampling syringe



## Flushing the system

- When flushing the system, a spare sample bottle is usually repeatedly filled and emptied into the waste. It is good practice to measure the oil temperature using the last bottle that will be discarded, as this avoids having to put the thermometer into the actual sample.

## Taking the sample

- Wherever possible, try to take samples during times of relatively steady loads and temperature – in other words, when the equipment is at equilibrium. (This is particularly important with transformers, as if the sample happens to be taken after the transformer has cooled following a long period of running at full load, the breakdown voltage of the oil will be much lower than normal. This is because moisture in the paper insulation will have migrated to the oil during the period of full load, and will not yet have had time to migrate back. This is usually considered to be a normal phenomenon, but it is possible that it may also be a factor in so-called ‘sudden death’ transformer incidents where, for no apparent reason, a seemingly healthy transformer suddenly fails. This is another good reason for recording as much information about the transformer as possible and for trending results to look for unexplained changes).
- **Do not** take samples when it is raining or snowing, or when the relative humidity is above 50%, as there is a high probability that samples taken in these conditions will be contaminated.
- **Do not** take samples when it is windy, as dust blown by the wind may contaminate the sample.
- Try not to take samples when the ambient temperature is high, as perspiration is a common source of contamination problems.

## 9. Preparing the test vessel

Successful dielectric breakdown voltage testing depends not only on obtaining a good sample, as discussed in the previous section, but also on ensuring that the test vessel is properly prepared. The preparation of the test vessel can be divided into two key elements – the first is storing, cleaning and filling, and the second is setting the electrode gap.

### 9.1 Storing and cleaning test vessels

IEC 60156 recommends that a separate test vessel assembly is used for each type of insulating fluid that it is required to test. This standard requires that the test vessels are filled with dry insulating fluid of the type that they will be used to test, then covered and stored in a dry place. ASTM offers an alternative option of storing the vessels empty in dust-free cabinet.

Immediately prior to testing, vessels stored full must be drained and then all internal surfaces, including the electrodes, rinsed with fluid taken from the sample to be tested. The vessel should then be drained again, and carefully filled with the test sample, taking particular care to avoid the formation of bubbles.

If the vessel was stored empty, or if it is to be used for a different type of fluid from that with which it was filled during storage, it should be cleaned with an appropriate solvent before the rinsing and filling procedures described above are followed. ASTM D1816 specifies the use of a dry hydrocarbon solvent such as kerosene, which meets the requirements of D235. Solvents with a low boiling point should not be used as these evaporate rapidly, cooling the vessel and giving rise to the risk of condensation. Solvents commonly used include acetone and, in the USA, toluene. Toluene is banned in Europe.

Use lint-free clean-room wipes to clean the vessel. Do not use paper towels as they may introduce particles that hold moisture, causing breakdown values to be dramatically reduced. Touching the electrodes or the inside of the vessel should be avoided and during cleaning, the electrodes should be checked for pitting or scratches that may cause breakdown voltage values to be decreased.

### 9.2 Setting the electrode gap

Setting the electrode gap accurately is very important, as the results obtained are only valid if the gap is correct. A big problem is movement of the electrodes after the gap has been set and for this reason, many users of oil test sets check the electrode gap frequently – sometimes before every test. A better solution is to use test sets where the electrodes can be locked in position, such as the instruments in Megger's latest OTS range.

Megger recommends the use of flat, smooth gap gauges. The latest Megger gauges have a black anodized coating, which not only provides a smooth surface but also shows when the gauge is worn, as the shiny aluminium starts to show through the coating.

### 9.3 Hints and tips for vessel preparation

- If rinsing the test vessel with the sample oil before testing, it is most important to immediately fill the test vessel with the oil sample to be tested. Any significant delay will result in the oil film on the vessel's walls absorbing water from the air, and since the walls have a large surface area, this will contaminate the oil sample and reduce the breakdown voltage once it has been mixed with the sample.
- Pour the oil into the vessel swiftly with minimum turbulence so as not to entrap air.
- Allow the sample to stand for a few minutes before the testing to allow air bubbles to clear.
- Do not leave the sample in the vessel to stand for too long before testing as it will absorb water from the air in the headspace above it. This will reduce the breakdown voltage.
- If you are using an impeller stirrer that utilises a baffle plate to exclude air from the oil sample ensure that:
  - Oil does not pass over the upper surface of the baffle plate
  - Oil is in full contact with the underside of the baffle plate
- The use of a magnetic bead for IEC60156 will circulate oil in the lower portion of the test vessel, whereas the impeller will circulate all of the oil in the test vessel. The magnetic bead therefore has the advantage that moisture absorbed by oil in contact with air is not stirred into the sample, avoiding unwanted contamination.
- Remember that the rules of cleaning and preparing the vessel also apply to the magnetic bead, impeller, baffle plate and electrodes, not just the vessel walls.
- When performing continuous testing of many oil samples, such as in laboratory environments it is important to clean or rinse the test vessel between every sample tested.
- Always refer to the appropriate test standard to ensure the preparation is performed as specified.

## 10. Why measure oil sample temperature?

The breakdown voltage of an oil sample increases significantly with temperature. For example, a natural ester sample with a breakdown voltage of around 35 kV at 30 °C could easily have a breakdown voltage of nearly 60 kV at 70 °C. For this reason, all oil test standards specify that the temperature of the sample must be recorded in the test report. Here, from the chart included earlier in this technical guide, is the section that refers to temperature.

Standards		ASTM D 1816	ASTM D 877		IEC 60156
			Procedure A	Procedure B	
Laboratory test temperature	Liquid	At ambient - must record	20-30 °C must record temp as collected and when tested	20-30 °C must record temp as collected and when tested	15 - 20 °C for referee tests
	ambient	20 - 30 °C	Must record	Must record	Within 5 °C of oil sample
Outside test temperature	Liquid	At ambient - must record	Must record	Must record	15 - 25 °C
	ambient		Must record	Must record	Within 5 °C of oil sample

**Note** The trending of test results to identify changes in breakdown voltage is only valid if the sample and ambient temperatures for all results have been taken into account. Some breakdown testers measure oil temperature automatically. This helps to ensure that the sample temperature has been measured and avoids the possibility of introducing contamination by placing a thermometer into the oil sample.

## 11. How do I know whether my test results are valid?

This extract from the chart comparing standards, which was included in full earlier in this application note, shows that each standard specifies different conditions that must be met if the results of a test are to be accepted as valid.

Standards	ASTM D 1816	ASTM D 877		IEC 60156
		Procedure A	Procedure B	
Valid test conditions	If breakdown does not occur at 2 mm, reduce gap to 1 mm. Tests must be repeated if range of BD voltages recorded are more than 120% of mean with 1 mm electrode gap and 92% of mean with 2 mm electrode gap.	Tests must be repeated if range of BD voltages recorded are more than 92% of mean. If range of 10 BD voltages is more than 151% investigate why.		Expected range of standard deviation/mean ratio as a function of the mean provided as a chart.

- Mean is the average of the breakdown values recorded in the test sequence. For example, if the breakdown values are 33 kV, 37 kV, 32 kV, 35 kV, 38 kV and 34 kV, the mean value would be the total of these results – 209 – divided by the number of results – 6 – which gives a mean value of  $209/6 = 34.83$  kV. (Note that in this example there are six results as required by the IEC standard. The ASTM standards require either five or ten results.)
- Range of breakdown voltage is referred to in the ASTM standards. For example, D877 specifies that the test sequence must be repeated if the range of breakdown voltages recorded is more than 92% of their mean value. Two examples will make this easier to understand.
  - In the first example, the breakdown voltages recorded are 43, 45, 52, 40 and 38 kV. The lowest value is 40 kV and the highest is 52 kV, so the range is 12 kV. The mean of the recorded values is 43.6 kV, so the range is only  $12/43.6 \times 100\% = 27.5\%$  of the mean value. These test results are, therefore, valid.
  - In the second example, the breakdown voltages recorded are 33, 45, 52, 18 and 20 kV. The lowest value is 18 kV and the highest is 52 kV, so the range is 34 kV. The mean of the recorded values is 33.6 kV, so the range is  $34/33.6 \times 100\% = 101\%$ . This is above the 92% limit, which means that the test must be repeated.
- Standard deviation - In IEC 60156, there is a graphical representation of standard deviation – otherwise known as the coefficient of variation – versus the mean breakdown voltage. Calculation of the mean has already been covered, but what about the standard deviation? IEC 60156 does not explain how to calculate this. The procedure however, is to calculate the difference between each of the six test results and the mean value of those test results. Square each of the differences and add them together. Divide the figure obtained by 2, and then take the square root. The final answer is the standard deviation for the set of test results.

IEC 60156 states that, for the test results to be considered valid, the following procedure must be followed:

- Perform six tests
- Calculate the mean of the results
- Calculate the standard deviation (see above)
- Divide the standard deviation by the average value, noting that scatter is expected and acceptable. (See the chart at the end of IEC 60156)
- If the value is acceptable, conclude testing
- If not, perform six more tests
- Repeat the calculations using all 12 results

## 12. How do I know if my fluid has passed the test?

The manufacturer of the insulating fluid normally quotes typical breakdown values for both new and in-service fluid in its data sheets. In addition, the test standards refer to oil condition standards that provide guidance about the acceptability of results.



## 12.1 USA – ASTM and IEEE standards

**D877** as already mentioned, is usually only recommended for the acceptance of new oil from a supplier. However, some oil testing laboratories still recommend its use for specific in-service applications. In these cases, a breakdown voltage of 30 kV or more is usually considered to be acceptable, with values below 25 kV unacceptable. Values between 25 and 30 kV are considered questionable. For new oil, a minimum value of 30 kV is normally specified.

Typical breakdown values using D877 test methods	
Oil type	New oil
Mineral oil	45 kV
Silicone oil	40 kV
HMWM	52 kV
Sythetic ester	43 kV
Natural ester	56 kV

**D1816** is more widely used and is accepted by the IEEE as the test method to be used for dielectric breakdown testing for the acceptance and maintenance of insulating oil. The IEEE C57.106 standard incorporates the D1816 limits – which are shown below – for new and in-service oil. Note that the values provided in this table are for mineral oil.

### IEEE C57.106-2006

#### IEEE Guide for acceptance and maintenance of insulating oil in equipment

Applications	Voltages class/group	D1816 (1 mm gap)	D1816 (2 mm gap)
New mineral insulating oil as received from supplier	Not specified	>20 kV	>35 kV
New mineral insulating oil received in new equipment, prior to energisation	≥69 kV	>25 kV	>45 kV
	69 to 230 kV	>30 kV	>52 kV
New mineral insulating oil - processed from equipment, prior to energisation	230 to 345 kV	>32 kV	>55 kV
	≥345 kV	>35 kV	>60 kV
Service aged insulating oil - for continued use (Group 1)	≥69 kV	>23 kV	>40 kV
	69 to 230 kV	>28 kV	>47 kV
	≥230 kV	>30 kV	>50 kV
Shipments if new mineral insulating oils OCB (Oil Circuit Breaker)	OCB	>20 kV	>30 kV
New OCB insulating oil - after processing, prior to energisation	OCB	>30 kV	>60 kV
Service aged OCB insulating oil - for continued use	OCB	>20 kV	>27 kV
New mineral oil for LTC (Load Tap Changer), prior to energisation	LTC	>35 kV	>55 kV
Service aged LTC insulating oil - for continued use	LTC - Neutral	>20 kV	>27 kV
	LTC - ≥69 kV	>25 kV	>35 kV
	LTC - >69 kV	>28 kV	>45 kV

## 12.2 International IEC standards

**IEC 60156** uses acceptance values that are contained in two further standards, IEC 60296 and IEC 60422.

**IEC 60296** – Fluids for electrotechnical applications – Unused mineral insulating oils for transformers and switchgear. As its title indicates, this standard applies only to new, unused oil as received from the manufacturer, which must have a dielectric breakdown voltage of 30 kV or more, determined using the IEC 60156 test method. Oil that has been vacuum filtered in a laboratory must have a minimum dielectric breakdown voltage of 70 kV.

**IEC 60422** – Mineral insulating oils in electrical equipment – Supervision and maintenance guide. This standard prescribes acceptable dielectric breakdown values for new oil (after filling but before energising) and for in-service oil. The values are:

New oil	
Equipment voltage	Dielectric BD voltage
≥72.5 kV	>55 kV
≥72.5 kV ≥170 kV	>60 kV
>270 kV	>60 kV


In-service oil	Dielectric BD voltage		
Equipment voltage	Good	Fair	Poor
≥72.5 kV	>40 kV	30 - 40 kV	>30 kV
≥72.5 kV ≥170 kV	>50 kV	40 - 50 kV	>30 kV
>270 kV	>60 kV	50 - 60 kV	>50 kV

The IEC recommends that if values are in the ‘fair’ range, testing should be performed more frequently, and that the test results should be crosschecked with other testing methods. If the test results are in the ‘poor’ range, the oil must be brought back into a good state by reconditioning. This might, for example, involve filtering and drying the oil.

## 13. Can new oil fail a dielectric breakdown test?

The simple answer is yes, new oil can fail a breakdown test. Sometimes users suspect that their test set is faulty because it is failing new oil. When the test set is checked however, almost invariably no fault is found. The true situation is that the oil being tested really does have a low breakdown value, so the test results are actually correct, even if they are hard to accept.

The problem usually relates to the way the oil has been stored. Often oil is delivered in drums that are stacked on site in readiness for filling a transformer or other equipment. In hot climates, pressure rises in the drums when they are exposed to the heat of the sun, but falls at night when the ambient temperature drops. Sometimes this pressure cycling damages the seals on the drums, and they start to allow the penetration of moisture. It only takes a very small amount of moisture to degrade the properties of the oil to the extent that it fails the test.



To give an idea how quickly oil can deteriorate, after a drum of new oil is opened it is often less than three weeks before it fails a breakdown test.

In circumstances where such problems may occur, it is useful to verify the performance of the oil test set, so that the real situation can be quickly determined. One approach is to use a voltage check meter to ensure that the test set is supplying the correct voltage. An alternative method is to perform a breakdown test in air with the instrument electrode gap set as wide as possible. If the value obtained when the instrument is new or when it has recently been calibrated is recorded, this can be used as a reference value to confirm that it is still working correctly at a later date.

Note that some variation in the air breakdown test results are to be expected, due to changes in atmospheric pressure and humidity, but this is a good 'quick check' that has the advantage of requiring no additional equipment. As a guide, Megger's older OTS AF2 instruments typically give a breakdown voltage of around 37 kV in air, while instruments in the newer OTS AF series typically give a result of approximately 19 kV.

### 14. Can I verify my test instrument performance by comparing the results from two instruments on the same sample of oil?

Yes, but you must consider these potential issues:-

- Even if you can transfer the oil sample from one instrument to another in the same test vessel you may not get exactly the same test results. This is because by-products produced by the breakdown arcs from the first set of tests are now somewhere in the oil sample and likely to effect any subsequent tests.
- Changes in oil sample temperature will produce changes in test results.
- Pouring oil from one vessel to another vessel is likely to produce bubbles that will introduce additional air and moisture to the sample resulting in lower test results than the previous testing
- If the time between comparative tests on the same oil sample is excessive it will absorb moisture from the air, even if a test vessel has a lid as it is not sealed in the same manner as an oil sample bottle.
- Two oil samples taken from the same drum are in effect two different samples of oil. They can, if particle contamination is present, exhibit different levels of breakdown voltage.
- Oil with unknown contamination can cause large changes in test results, particles or moisture can circulate in the oil between the electrodes, causing variation in test results. This can easily lead to an incorrect conclusion with regard to the performance of the test instruments being compared. That is why five or six breakdown tests are performed during each test sequence to take into account this variation
- Cleanliness is extremely important to ensure reliable test results that can be reliably compared.

### 15. How do oil test sets detect dielectric breakdown?

Efficient breakdown detection is very important as it determines the ability of the instrument to correctly register the breakdown voltage of the oil, and also to limit the amount of energy dissipated in the arc in the oil. For these reasons, the oil testing standards specify the precise conditions that must be met to trigger the instrument's breakdown detection system. There are, however, differences between the standards, as shown in the table below:

Standards		ASTM D 1816	ATSTM D 877		IEC 60156
			procedure A	procedure B	
Breakdown	Definition	When output voltage had dropped to below 100 V	When output voltage had dropped to below 100 V	When output voltage had dropped to below 100 V	When output current has been 4 mA or more for at least 5 ms
Test voltage switch off time following breakdown	Normal (e.g. mineral oil)	Not specified	Not specified	Not specified	<10 ms
	Silicon oil	Not specified	Not specified	Not specified	<1 ms

Modern test instruments detect breakdowns directly on the instrument output, which improves detection accuracy. Very few instruments are capable of breakdown detection that meets both IEC and ASTM requirements. Instruments in the new Megger OTS range are an exception as they can detect breakdown in terms of current increase or in terms of voltage drop.

It is worth noting that some older instruments do not include automatic breakdown detection. IEC 60156 states that the circuit applying the test voltage may be opened manually if a transient spark (audible or visible) occurs between the electrodes. This statement is included specifically to cover the use of these older instruments where, if a spark is heard or seen between the electrodes, a breakdown has occurred and the test is complete.

## 15.1 Testing silicone oil

Silicon oil has much higher viscosity than mineral oil, and is more susceptible to the breakdown arc producing carbon track solids. For these reasons, the standards specify special conditions when testing silicone oils:

**IEC 60156** reduces the test voltage switch-off time after breakdown detection from the standard 10 ms to 1 ms for silicone oils.

**ASTM D877** and **D1816** both specify that, for silicone oils, the energy discharged into the breakdown arc must not exceed 20 mJ.

## 16. Choosing an oil dielectric breakdown voltage test set

It would be easy to assume that because all dielectric breakdown voltage test sets have to work to the same strict test standards, there is little to choose between them. This is far from true and a little effort put into selecting the best instrument will be amply repaid by time and cost savings over the life of the instrument.

The first decision to make is whether the tests will be performed in a laboratory or on-site, as there are differences between laboratory and portable instruments.

### 16.1 On-site versus laboratory testing

The relative merits of on-site and laboratory testing have long been debated by engineers and the debate is still going on. The underlying issue is that test results are greatly influenced by contamination. There are those who argue that this means it is better to test on-site because, if a sample has to be bottled and the bottle sent to a laboratory, there will always be doubt about whether the bottle was cleaned adequately before use and whether it was sealed sufficiently well to guard against contamination in transit.

Others will point out that testing on site provides no guarantees against contamination, as the most likely time for a sample to be contaminated is while it is being collected. They may also argue that, if a sample is sent to a laboratory, it will be tested by a skilled technician who will follow all of the procedures needed to ensure accurate results. Conversely, tests carried out on site are performed under less-than-ideal conditions and often under significant time pressure, which may lead to errors.

On-site testing undeniably has the benefit of immediacy. If a suspicious result is obtained, it is usually possible to repeat the test without delay and, if the problem is confirmed, the affected transformer can, at least in principle, be taken out of service immediately.

There is no clear answer to the question of whether on-site or laboratory testing is best. It is really a matter for individual users to decide which approach best suits their particular circumstances. Megger is always happy to provide advice in making this decision and, as a company that manufactures both on-site and laboratory test sets, the impartiality of that advice is guaranteed.

### 16.2 General instrument selection considerations

An instrument is a long-term investment, so it's important to choose one that's built to last and that is as future proof as possible. This section looks at additional considerations that apply to all oil test sets, while sections 15.3 and 15.4 deal with considerations relating specifically to laboratory instruments and portable instruments respectively.

Instruments in Megger's new OTS range have been designed to address all of the issues identified in these sections.

## 16.2.1 Cleanliness

- **Insulating fluids are hazardous.** Skin contact with these fluids should be avoided. Instrument should be designed to minimise the risk of spillage, and to make spillage as easy as possible to deal with if it should occur.



- **Test vessel ease of cleaning.** The vessel should be designed so that it has no corners, as these trap dirt and are difficult to clean.



- **Chemical resistance.** The test vessel should be made of a material that is highly resistant to chemical attack. Glass is not the best choice because of its fragility. A correctly specified moulded material will not affect test results – note that even glass vessels have moulded bushes to support the electrodes, and these bushes are in contact with the test sample.


The test vessels on the new Megger OTS range instruments are manufactured from a material which is virtually shatterproof and which is widely used in the chemical industry because of its excellent resistance to chemical attack. The vessels have been extensively tested to confirm that they are unaffected by any of the chemicals likely to be encountered in oil testing applications.

- **Ease of pouring.** Always look for a test vessel that has a spout to aid pouring.



- **Testing to ASTM D1816.** Remember that this standard requires the oil sample to be stirred with an impeller, and also specifies that the test vessel must have a cover of be fitted with a baffle to prevent air from coming into contact with the sample. In many instruments, oil is displaced by the lid, resulting in messy spills. Look for a test set that does not have this problem.


The stirrer lid for the new Megger test vessel has a unique baffle plate design. When the test vessel is filled with oil to the fill line and the stirrer lid fitted, the baffle plate touches the oil without displacing it. No oil is spilled, so no clean up is necessary.



- **Test chamber ease of cleaning.** Like the test vessel itself, the test chamber should be designed so that it has no corners to trap dirt and contaminants.



- **Dealing with spillage.** Inevitably, oil is sometimes spilled into the instrument's test chamber. With many instruments, the only way of removing this spilt oil is by using cloth or tissues to mop it up, a task that is both time consuming and inconvenient. Make life easier by specifying an instrument that has a test chamber with a means of quickly and easily draining spilled oil.




## 16.2.2 Ease of use


- An instrument that is difficult to use is not just an irritation, it also wastes time and reduces productivity. Look for an instrument that has a clear and straightforward user interface. Operation of the instrument should be intuitive and it should not be necessary to repeatedly refer to the user guide. This is particularly important for instruments that are made available for hire, and in cases where they will be used by engineers and technicians who do not perform oil tests regularly. Here are some specific points to consider.
- **User interface.** The user should be able to easily identify and access the main functions of the instrument, such as test method selection, test sequence set up, stored data functions and user settings including, for example, the operating language.

The user interface in the new Megger OTS AF instruments has a clear and easy to understand menu system, with simple navigation keys. The interface presents the user with a set of six tabbed top-level windows, as shown below. The left and right arrow keys on the keypad navigate between the tabs.


## Home – Test set up




## Set up – Instrument settings




## File – data functions




## Information – help




## Tools – additional functions



## Language - selection



- **Display.** The display should be clear and bright, and preferably in colour. The display is a window into the functioning of the instrument – the bigger and clearer this window is, the better.
- **Test chamber access.** Good test chamber access will aid instrument operation and cleaning. Deep and dark chambers are bad, not least because they make it difficult to see if the chamber is dirty. Most users prefer top access instruments.
- **Automatic testing.** Most modern instruments support automatic testing, and will follow the test sequence specified in the chosen test standard with little or no operator intervention. Low-cost manual instruments are still available, but think carefully about day-to-day operation is needed before purchasing one of these. They may be appropriate for some users, but a high degree of skill is needed to use them successfully.



### 16.2.3 Ownership costs

- **Additional and replacement test vessels.** As mentioned earlier, IEC 60156 recommends that a separate test vessel is used for each type of fluid to be tested. Many laboratories adhere to this advice and as a result, have six or more test vessels. It is important, therefore, to check the price of spare vessels before purchasing an instrument. Some are very expensive and purchasing several vessels can greatly increase overall costs.
- **Broken test vessels.** The previous comments about the price of test vessels are also relevant to broken test vessels, but note that moulded test vessels are much less fragile and far less likely to break than glass vessels. When correctly specified a moulded vessel will provide all the advantages of a glass vessel and more, at a lower price.



## 16.2.4 Sample temperature measurement

An instrument that features built-in automatic oil-sample temperature measurement offers important benefits. With these instruments, the temperature sensor is in contact with a thin section of the test cell; unlike a conventional thermometer, the sensor does not come into direct contact with the sample, and the risk of contamination is eliminated. Also the user is less likely to be tempted to save time by not waiting for the sample to cool before testing. Finally, there is no fragile thermometer to break.



Chamber mounted temperature sensor on the new Megger OTS range

## 16.2.5 Test voltage verification

- **Checking between calibration dates.** Test results are only as accurate as the instrument's ability to measure and control its output voltage, so it is good practice to verify the operation of the instrument between calibrations. This is not because instruments are likely to lose accuracy, but because if something occurs that affects calibration and it isn't noticed, it could mean that a lot of inconvenient and potentially expensive retesting is needed.
- **Voltage check meters.** Many manufacturers offer a voltage check meter that can be fitted to the instrument in place of the measuring vessel. This allows the user to compare the voltage shown on the check meter with that shown on the instrument display. Check meters are not sufficiently accurate to use as a calibration standard, but they provide a very good way of detecting changes in instrument calibration. Users should record the check meter readings each time a voltage check is carried out, so that changes can be easily identified. If any significant change is detected, the instrument should not be used until it has been returned to the manufacturer for servicing and recalibration.

The Megger VCM100D check meter can be used with any of the portable or laboratory test sets in the new OTS range. Since the meter is not oil filled, there is no risk of leakage. It also features a bright, easy-to-read LED digital display.



## 16.2.6 Changes to test standards

- When choosing a test instrument, it is essential to be sure that it can handle future changes to test standards and the introduction of new test standards. It should be possible to easily load new test sequences into the instrument to reflect these changes. It is worth checking that the instrument supports custom test sequences that can be programmed by the user.

Testers in the Megger OTS range allow existing pre-loaded standards to be updated and new standards to be added via a standard USB flash drive. New and amended standards are made available for download from the Megger website.





## 16.2.7 Reliability of test results

**Note** There are five key areas that directly affect an instrument's ability to produce accurate test results. These are cleanliness, the method of breakdown detection, the speed at which the instrument switches off the test voltage after a breakdown, the accuracy of the voltage measurement and the accuracy of the electrode gap setting mechanism. The issues of cleanliness and output voltage verification have already been discussed; the remaining issues are dealt with here.

- **Breakdown detection.** Ensure that the instrument complies with the requirements of the test standard(s) that will be used. Megger OTS AF test sets offer both voltage- and current-based breakdown detection, to provide compliance with ASTM and IEC requirements.
- **Voltage switch-off time.** The faster the voltage is switched off after breakdown, the better. This is because arcing continues in the oil until the voltage is switched off, and arcing results in decomposition of the oil to produce by-products that can affect subsequent test results. For this reason, the IEC 60156 standard requires that the test voltage is switched off in less than 10 ms for mineral oil and in less than 1 ms for silicone oil. The ASTM standards require that no more than 20 mJ of energy is discharged into silicone oil, and a fast switch-off time is needed to achieve this.



The switch-off time for instruments in the new Megger OTS range is typically 10  $\mu$ s - 100 times faster than required to minimise decomposition of silicone oil.

- **Electrode gap setting.** An accurate electrode gap is extremely important for obtaining accurate test results. If the gap is too large, the test result will be too high and conversely if it is too small, the test result will be too low. The easier it is for users to adjust the gap, the more likely it is that it will be accurately set.
- **Ease of gap adjustment.** The most popular method for adjusting the electrode gap is with a thumbwheel. With this method, no tools are needed and it is easy to adjust the gap.
- **Unwanted movement of electrodes.** A common problem is that the electrodes move while the vessel is being handled or cleaned, resulting in an inaccurate gap setting. This is particularly true of instruments that have thumbwheel electrode gap adjustment as thumbwheels are easy to move accidentally. In addition, backlash in the adjustment mechanism can make movement of the electrodes inconsistent, which means that they are difficult to adjust accurately, and marked graduations are rendered inaccurate. Because of this, many test set users resort to checking the electrode gap every day or even before every test. Some older instruments provide screw clamp locking that is effective in preventing electrode movement, but it is hard to see and use.

Megger's new OTS instruments have convenient thumbwheel adjustment for setting the electrode gap, but have the unique benefits of pre-tensioned electrodes and an easy-to-use locking mechanism that ensures the electrodes can be adjusted accurately and never move accidentally once set.

### Precision thumbwheel gap adjustment

### Gap is then locked with unique mechanism



## 16.3 Selection considerations specific to laboratory instruments

Users of oil test sets who work in the laboratory have priorities that are different from those of users who work in the field. For example, in a laboratory environment, the weight of the instrument is usually not a concern, but the ability to carry out large numbers of tests rapidly in succession most certainly is. For this reason, Megger produces separate ranges of dielectric breakdown test sets for laboratory and field applications. This section deals specifically with laboratory instruments.

### Test chamber access

When large numbers of tests are being performed, easy access to the test chamber is essential. For this reason, instruments in the new Megger OTS AF range have large test chambers that are fast to clean and convenient to use.

OTS PB chamber



OTS AF chamber



### Test result recording

Recording test results and producing reports are among the most time-consuming tasks in a laboratory. Speed and ease of data entry are therefore, key considerations when choosing a test instrument. It is a matter of personal preference whether the user favours entering data into a PC or directly into the instrument, but in either case the task should be fast and straightforward. If a laboratory instrument management system (LIMS) is in use, it is important to check that the test set can output data in a compatible format.

Megger OTS AF laboratory instruments have a keypad that is specifically designed to speed up data entry, and all current OTS instruments allow data to be transferred to a PC using a standard USB flash drive. In addition, OTS instruments are supplied with a free copy of Megger's PowerDB Lite software, which allows professional test reports to be generated quickly and easily, and can export test data in Excel and CSV formats. The instruments are compatible with the full version of PowerDB.

OTS AF keypad



## Barcode scanning

A timesaving feature that is often worth considering is barcode scanning, as it provides a fast and error-free way of entering sample identification numbers.

### OTS AF instruments can scan barcode labels using the optional barcode scanner



## 16.4 Selection considerations specific to portable instruments

The priorities of those who use portable instruments are different in several key areas from those of laboratory instrument users. Among the main points to consider for portable instruments are:

### 16.4.1 Weight

- Portable instruments routinely have to be carried from a van to a transformer. Weight is of particular interest, especially when the instrument has to be carried by just one person. The government or the organisation you work for may have rules on the maximum weight that you should carry; often it is specified at 25 kg. From the user's point of view however, the lighter the instrument the better.

Megger has made great efforts to minimise the weight of its latest OTS instruments and, starting at just 15.4 kg, the OTS60PB oil dielectric breakdown test set is well within most health and safety guidelines on weight.

### 16.4.2 Ease of carrying


How easy the instrument is to carry is an important consideration. Handles on the sides of the instrument are generally easiest to use, as they provide a more natural lifting position than a single handle on the top of the instrument.



### 16.4.3 Toughness

When an instrument is damaged, testing must stop until it can be repaired. Portable instruments are particularly likely to take a few knocks during transportation and use, and they must be able to withstand this treatment. Items might fall on the instrument when it is in the back of a van, and it is desirable to choose an instrument that has been designed with this type of incident in mind.

Instruments in the new Megger OTS range have curved tops that deflect falling objects. In addition, the chamber cover and display protection window are both manufactured from strong impact-resistant transparent polycarbonate material. Not only is the chamber cover very strong, it also allows the discharge to be seen clearly.



Occasionally instruments also get dropped. The most frequent types of damage suffered by oil test sets dropped from a modest height are crumpled corners and/or twisting of the instrument frame. Dielectric breakdown test sets include a heavy transformer that must be properly supported, and a quick look at the underside of the instrument will often give a good idea of whether it has sufficient structural strength. In particular, the corners should be checked to see whether they are likely to be able to withstand impacts. It is also worth bearing in mind that lighter instruments are less likely to suffer damage.

## Megger OTS rubber feet / corner protection



### 16.4.4 Transferring stored data

When the time comes to transfer stored test data from the instrument to a PC, consider the inconvenience of having to carry the instrument into an office so that it can be connected to the PC. A better solution is to choose an instrument that can store test results on a standard USB flash drive, as then it is only necessary to carry the USB flash drive into the office.

All of the new Megger OTS instruments transfer data via a USB flash drive. This method is simple, convenient and secure.



### 16.4.5 Power sources

Many portable instruments can operate only from their internal battery. Bearing in mind that the typical time needed to fully recharge the battery is around four hours, this can be inconvenient if the instrument is needed urgently and is found to have a discharged battery, or if a large number of tests have to be carried out during a fixed shutdown period. It is better to choose an instrument that can operate either from its internal battery or direct from the mains.

Instruments in the new Megger portable range can operate from either a mains supply or from their internal batteries. The batteries can be charged from the mains or from a 12 V dc vehicle socket, using the 12 V charger lead supplied as standard.

		Megger NEW range					Megger
		Laboratory optimised instruments			Portable optimised instruments		
Feature	Sub-feature	OTS60AF	OTS80AF	OTS100AF	OTS60PB	OTS80PB	OTS60SX
Test voltage etc	Max. output voltage	60 kV	80 kV	100 kV	60 kV	80 kV	60 kV
	Direct output voltage measurement	■	■	■	■	■	
	Voltage slow rate (kV/s)	0.1 to 5 kV/s	0.1 to 5 kV/s	0.1 to 5 kV/s	0.1 to 5 kV/s	0.1 to 5 kV/s	0.5, 2 and 3 kV
	Switch off time on flashover	< 10 µs	< 10 µs	< 10 µs	< 10 µs	< 10 µs	not specified
	Direct output breakdown detection	■	■	■	■	■	
	Voltage (ASTM) and current (IEC) triggered breakdown detection	■	■	■	■	■	
	Measurement accuracy ±1 kV or better	■	■	■	■	■	■
Measurement accuracy ±0.1 kV or better	■	■	■	■	■	■	
Oil temperature measurement	Built in to instrument	■	■	■	■	■	
Test vessel	Moulded - impact resistant	■	■	■	■	■	■
	Magnetic stirring	■	■	■	■	■	
	Propeller stirring	Option (CTB) *	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option
	Auto power connection for above	■	■	■	■	■	
	Spout for pouring	■	■	■	■	■	■
Test chamber	Spilt oil drain	■	■	■	■	■	
Electrode gap adjustment	Micrometer style adjustment	■	■	■	■	■	
	Lockable as above	■	■	■	■	■	
Power supply	Mains operation	■	■	■	■	■	■
	NiMH battery				Option (CTB)*	Option (CTB)*	
	NiMH battery life				10 hours	10 hours	
	Lead acid					Option (CTB)*	
	Vehicle 12 V socket operation				■	■	
User interface	Multi language	■	■	■	■	■	Manual
	Memo entry	■	■	■	■	■	
	Multikey keypad	■	■	■	■	■	
	Display backlight	■	■	■	■	■	LED
	Colour TFT display	■	■	■	■	■	
Auto test sequences	ASTM	■	■	■	■	■	Manual
	IEC - and related	■	■	■	■	■	Manual
	Custom programmable	■	■	■	■	■	Manual
Types of oil tested	Includes silicone	■	■	■	■	■	
	Recording results int	■	■	■	■	■	
	Recording results USB flash drive	■	■	■	■	■	
	PowerDB Lite supplied	■	■	■	TBA	TBA	
Other data	Date and time on unit	■	■	■	■	■	
Printer	Built in	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option (CTB)*	
	External	Option	Option	Option	Option	Option	

\* CTB = configure to build

		Megger NEW range						Megger
		Laboratory optimised instruments			Portable optimised instruments			
Feature	Sub-feature	OTS60AF	OTS80AF	OTS100AF	OTS60PB	OTS80PB	OTS60SX	
Environmental	Operating temperatures	0 to +50 °C	0 to +50 °C	0 to +50 °C	0 to +50 °C	0 to +50 °C	0 to +40 °C	
	Storage temperatures	-30 to +65 °C	-30 to +65 °C	-30 to +65 °C	-30 to +65 °C	-30 to +65 °C	-40 to +70 °C	
Physical	Dimensions (W x H x D) m	0.58 x 0.42 x .029	0.58 x 0.42 x .029	0.58 x 0.42 x .029	0.52 x 0.34 x .025	0.52 x 0.38 x .025	0.4 x 0.33 x .024	
	Weight without test vessel (min. to max. depending on configured items)	29.58 kg to 30.04 kg	29.58 kg to 30.04 kg	29.58 kg to 30.04 kg	15.38 kg to 17.94 kg	19.48 kg to 25.54 kg	17.5 kg	
	Weight with test vessel (min. to max. depending on configured items)	30.68 kg to 31.14 kg	30.68 kg to 31.14 kg	30.68 kg to 31.14 kg	16.48 kg to 19.04 kg	20.58 kg to 25.64 kg	not specified	
Accessories supplied	Protective carry case				Option (CTB)*	Option (CTB)*	Option	
	Test vessel	■	■	■	■	■	Option	
	Gap gauges	■	■	■	■	■	■	
	Electrode kit (IEC, ASTM or International)	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option	
	Mains/line supply lead with local plug	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option (CTB)*	Option (CTB)*		
	Transport case				Option	Option		
Voltage check meter available	Digital display	■	■	■	■	■		

\* CTB = configure to build

We hope that the information in this technical guide covers the most common questions relating to oil dielectric breakdown voltage testing. You can obtain additional advice from the Megger Technical Support Group (TSG) by calling +44 (0) 1304 502102, or by sending an email to [dovapptechtsupp@megger.com](mailto:dovapptechtsupp@megger.com)

You can find full details of the latest Megger oil dielectric breakdown voltage test sets at [www.megger.com](http://www.megger.com).

Megger reserves the right to change product specifications without notice.

## Important Warning!

This technical guide is intended to provide general guidance and advice on carrying out dielectric breakdown voltage tests and explains the important advantages and user benefits provided by Megger OTS test sets. It does not remove the necessity for reading and understanding the appropriate standard or standards when carrying out dielectric breakdown voltage tests.









Australia  
Megger Pty Limited  
Unit 1, 11-21 Underwood Road  
Homebush  
NSW 2140  
T: +61 (0)2 9397 5900  
F: +61 (0)2 9397 5911

Canada  
Megger Limited  
Unit 106-550 Alden Road  
Markham, Ontario L3R6A8  
Canada  
T: 1 416 298 6770  
F: 1 416 298 0848

France  
23 rue Eugène Henaff  
ZA du Buisson de la Couldre  
78190 TRAPPES  
T: 01 30 16 08 90  
F: 01 34 61 23 77

Germany  
Megger GmbH  
Obere Zeil 2  
61440 Oberursel  
Deutschland  
T: 06171-92987-0  
F: 06171-92987-19

India  
Megger (India) PVT Limited  
211 Crystal Paradise Mall  
Off Veera Desai Road  
Andheri (W)  
Mumbai - 400 053  
T: +91 22 2674 0468  
F: +91 22 2674 0465

Spain  
Oficina Central:  
Calle la Florida 1,  
Nave 16,  
Parque Empresarial Villapark  
28670 Villaviciosa de Odón,  
Madrid  
T: +34 916 16 54 96  
F: +34 916 16 57 84

South Africa  
Megger  
PostNet Suite 242,  
Private Bag X3,  
Bloubrigrant, 7443  
T: +27 (021) 557 6572  
F: +27 (086) 521 4576

Sweden  
Megger AB  
Box 724  
182 17 Danderyd  
T: 08 510 195 00  
F: 08 510 195 95  
seinfo@megger.com

Switzerland  
Megger AG  
Felsweg 1  
5727 Oberkulm  
Aargau  
T: +41 62 768 20 30  
F: +41 62 768 20 33

United Arab Emirates  
Megger Limited  
PO Box 500503 @D/C13  
Office 209, Bldg 14,  
Dubai Internet City,  
UAE  
Te: +971 4 443 5489

United Kingdom  
Archcliffe Road  
Dover  
Kent CT17 9EN  
T: 01304 502101  
F: 01304 207342

USA Dallas  
Megger Limited  
4271 Bronze Way  
Dallas  
Texas  
75237-1088  
T: +1-214-333-3201  
F: +1-214-331-7399

USA College Station  
Megger  
4064 State Highway 6 South  
College Station, TX 77845 USA  
T: 1-979-690-7925  
F: 1-979-690-0276

USA Valley Forge  
Megger  
Valley Forge Corporate Centre  
2621 Van Buren Avenue  
Norristown, PA 19403  
T: 1-610 676 8500  
F: 1-610-676-8610 (from the US 8610)

Megger Hong Kong Limited  
Workshop No. 8, 26/F,  
Mega Trade Centre,  
1 - 6 Mei Wan St., Tsuen Wan, N.T.,  
Hong Kong  
E: meggerhongkong@megger.com  
T: +852 2618 9964  
F: +852 2618 9001



Price £17.50 €20.00 \$25.00

# Megger®

WWW.MEGGER.COM