



Rhythm Turn

Low Air Loss and Dynamic Turning Mattress Replacement System



User Manual

Prius Healthcare USA
4027 Tampa Road, #3200
Oldsmar, FL 34677, USA

TEL: (813)854-5464
FAX: (813)854-5442

Content

1. The Purpose of this Manual	1
2. Product Description.....	1
Master Control Unit Features	
Mattress Features	
3. Technical Data	2
Master Control Unit	
Rhythm Turn Mattress Replacement	
Symbol Definition	
4. Instruction for Proper Use.....	3
5. Cleaning.....	6
The Mattress	
The Master Control Unit	
Replace Air Filter	
Waste Disposal	
6. Storage and Care	7
7. Maintenance and Troubleshooting.....	8
8. EMC Related Notification	9
9. Warranty.....	12

Warning

- ❖ Connect the Master Control unit to a proper power source.
- ❖ Don't use the system in the presence of any flammable gases. (such as Anesthetic Agents)
- ❖ Keep the pump and mattress away from source of liquid or open flame.
- ❖ Keep the mattress away from sharp objects
- ❖ The device is not AP/APG protected.
- ❖ Don't place the heating device close to the mattress system.

Caution

- ❖ Use the mattress under the physician's instruction.
- ❖ Re-positioning the patient in certain periods of time is still necessary when using this system.
- ❖ The Control unit can only be repaired by an authorized distributor. (The circuit diagram, repairable component parts list, and service manual are released only to authorized distributors).
- ❖ Don't drop the control unit.
- ❖ Do not store the system in direct sunlight or extreme cold conditions

1. The Purpose of this Manual

This operation manual is mainly focused on the set up, cleaning and routine maintenance of the Rhythm Turn-Function Air Therapy Support System. We recommend you keeping this manual handy to answer most of the question related to the system.

2. Product Description

The Rhythm Turn system, operated blower unit, is a very unique innovation of a specialized mattress replacement. The system is primarily designed for at risk patients or step-down intensive care units. It features continuous lateral rotation therapy in two different degree (20 degree and 40 degree), which gently turns the patient from side to side to significantly lower the risk of infection, pneumonia and other pulmonary complications – illnesses that significantly add to patient care costs and length of stay.

Master Control Unit Features


- The Master Control Unit is user friendly designed and most of the functions are self explained.
- Rotation angle can be independently selected for 20 degree or 40 degree.
- Rotation time can be adjusted in 5 min increments to 99 min. Or the caregiver can even select the Static Function that will cease the Rotation Function and provide only the True Low Air Loss Therapy.
- Auto Firm Function provides a uniform firmness for nursing procedure.
- Power failures produce an audio alarm for added safety.
- 10 digital scales of Soft/Firm Comfort Control.
- Double insulation to provide max silent operation.
- Foot board mounting rack provides the convenience of placement.

Mattress Features

- Individual air cushion design for maximum pressure distribution.
- Each air cushion has orifices to provide true Low Air Loss therapy.
- LAL Turning Mattress Replacement, eliminating the compromising effects of an existing mattress.
- Permanent inflated bed rails for added safety.

3. Technical Data







Master Control Unit

Model Name	Rhythm Turn Control Unit
Model No.	FC-PHR0026
Size	17.7”(L) x 6.8”(W) x 10.8”(H)
Weight	13.5 lbs
Dwell Time	3~95
Max Operating Pressure	50 mmHg
Rated Voltage	AC 110~120V
Rated Frequency	60 Hz
Fuse Rating	T5AH / 250V
Max Current	5A
Classification	Class I, Type BF Not AP or AGP type 
Operation Temperature	15°C ~ 35°C
Operation Humidity	30% ~75%
Mode of Operation	Continuous
Standard	IEC 60601-1, CAN/CSA C22.2 No. 601.1, IEC 60601-1-2

Rhythm turn Mattress Replacement

Model Name	Rhythm Turn Air Mattress
Model No.	FM-PHR0013
Size (cm)	91(W) x 203(L) x 20(H)
Weight (Kg)	20 Kg
Cells Material	TPU
Cover Material	AD with Quilting
Base Material	Nylon laminated PVC

Symbol Definition

	Refer To Instruction Manual / Booklet NOTE Background colour: blue Symbol : white		Waste Disposal
	Type BF Applied Part		Alternating Current
	Warning		Keep Dry

4. Instruction for Proper Use



1. Remove the existing mattress from the bed frame.
2. Replace the standard mattress with LAL Turning Mattress Replacement and make sure orient mattress so that the air tube is placed at the foot of the bed.
3. Secure the straps beneath the mattress to the bed frame.
4. Hang the Master Control Unit on the foot-board of the bed frame. Attach the air tubes connectors to socket on the left panel of the Master Control Unit. Be careful on the color matching between the connectors and socket. (black connectors to black socket, red connector to red socket)
5. Ensure the air hoses are not kinked under the mattress. (Could be verified by simple visual check) For detail Air hose connection please refer Explode Diagram.
6. Zip the low shear top cover to the mattress. The top cover should be loosely fit to the mattress.
7. Carefully plug power cord into a properly grounded power source. Turn on the master mechanical power switch on the right side panel. The STANDBY LED should illuminates.



8. Push the STANDBY/OPERATE button of the front panel. The OPERATE LED should now be lighted up and the Master Control Unit should now start to spin.



9. Push the AUTO FIRM button for fast inflation. Allow 4 ~ 7 minutes for full inflation. After the mattress is fully inflated, the caregiver can now transfer the patient on to the mattress. Push the AUTO FIRM again to release the fast inflation mode. (Note: The mattress can be inflated with patient lying on top).



10. Static Function: Push the static button and adjust the Contort Control by pressing the SOFT/FIRM button to achieve the maximum patient comfort. On this mode the system provides True Low Air Loss therapy. Perform a hand check by placing hand under the patient buttocks between cells and foam. The patient should have at least 4 cm of clearance between the buttocks and the bottom of the mattress.



11. Turning time can be adjusted by the CYCLE button. The time can be adjusted from 3 minutes to 95 minutes. (When Static Function is selected, the timing window would not show any digits).



12. The Master Control Unit is equipped with power failure alarm. With this function enabled, the Control unit generates a horn sound to signal to caregiver that main power failure. The alarm can be disabled by pushing the Alarm Reset Button on the front panel.



⚠ Caution: Immediate response by the operator is required with power failure alarm.


13. **LOCK-OUT:** The Master Control Unit is also equipped with a manual locking-out function. All function keys will be automatically disabled if the LOCK -OUT button has not been activated. When lock-out has been engaged, the “LOCK OUT” button will illuminate.

UNLOCKING


Unlocking the control panel is easy. Simply press the “LOCK OUT” button on the control panel for 3-5 seconds or recycle the power by turning off and on the main power switch.




14. 20 degrees turning function can be activated by pressing the “TURN” and select desire turning therapy. “LEFT TURN” allows the mattress to turn to left and back to horizontal. “RIGHT TURN” would have the same effect but turning to the right. The “FULL TURN” allows for full function of turning to left and right and should always activate with timer setting. The timer can be set by “CYCLE” button.



15. 40 degrees turning function can be activated by pressing the “BT” button and follow the operation instruction on step 14.



Mattress Turning Illustration



5. Cleaning

The Mattress

The mattress should be cleaned on the bed weekly using a damp soft cloth and mild detergent.

If top sheet (Top cover) or base (Bottom cover) becomes overly soiled, put on clean gloves, plastic gown and eye protection before removing top sheet or base and dispose of according to standard in function control procedures. Replace with clean covers.

Covers can be washed and thermally disinfected in a washing machine following below procedure: **(Never use phenol based cleaning solutions).**

Industrial cleaning

Break washes	cold	10 minutes
Main washes	60°C	6 minutes
Main washes	72°C	10 minutes
Extraction		2 minutes
3 Cold Rinses		
Extraction		5 minutes

Domestic cleaning

Pre-wash	cold	
Main Wash	72°C	10 minutes
Extraction		2 minutes
Cold Rinses		
Extraction		5 minutes

Tumble Drying or Tunnel Drying is not recommended.

Mattress Cells can be wiped over with a solution of sodium hypochlorite 1000ppm or any other non-phenolic germicidal solution.

The Master Control Unit

CAUTION

SWITCH OFF THE ELECTRICAL SUPPLY TO THE PUMP AND DISCONNECT THE POWER CORD FROM THE MAIN SUPPLY BEFORE CLEANING AND INSPECTION

The master control unit should also be cleaned weekly using a damp soft cloth and mild detergent.

The pump casing is manufactured from ABS plastic and if the case is soiled the pump can be wiped down with a sodium hypochlorite solution to dilution of 1000ppm or any EPA- approved hospital grade disinfectant. **(Don't use phenol base cleaning solution).**

The air filter should also be cleaned and checked as often as possible at a minimum of every six months. Air Filter can be removed by pinching center of the filter and pulling outward from the back of the Therapy control unit.

Replace Air Filter

1. Remove Air filter and Replace a new Filter.
2. Use a soft bristle brush to remove dust and difficult dried-on soil.



Waste Disposal

This Product has been supplied from an environmentally aware manufacturer that complies with the WEEE.

This product may contain substances that could be harmful to the environment if disposed of in places (landfills) that are not appropriate according the legislation. Please be environmentally responsible and recycle this product through your recycling facility at its end of life.



6. Storage and Care

Master Control Unit:

- Check the power cord and plug for abrasions or excessive wear.
- Plug in the unit and verify air flows from the units hose connection ports
- Place in plastic bag for storage.

Mattress Replacement:

- Check the air manifold for kinks or breaks. Replace if necessary.
- Pull out the CPR plug at the head of the mattress and disconnect the air feed tubes. All the air will now be expelled. Starting at the head end, the mattress can now be rolled. Use the base mounted straps for containment.
- Place in plastic bag of storage.

It is recommended the following guidelines are used whenever this system is being stored or transported another location:

Temperature limitations:	5°C ~ 60°C
Relative Humidity	30% ~ 75%

7. Maintenance & Troubleshooting

No daily maintenance is required. It is intended this equipment should only be serviced by properly qualified, authorized technical personnel. In case of minor trouble please refer as following Troubleshooting.

Symptom	Inspection Procedure	Possible Solution
Air is pumping out from the control unit but mattress is not inflating.	<ol style="list-style-type: none"> 1. Is the power source correct? Improper voltage may cause the pump to function abnormally and damage the control unit. 2. Is there any kinking tube? 3. Is there any air leakage from the air cells? 4. Is there any air leakage from air tube between mattress and control unit? 5. Has the air tube connector been connected properly? 	<ol style="list-style-type: none"> 1. Use power regulator. 2. Adjust the air tubes to enable smooth air flow. 3. Replace with new air cells 4. Replace with new air tubes 5. Re-connect the air tubes.
The Control Unit is not functioning.	<ol style="list-style-type: none"> 1. Check the power cord and the power voltage. 2. Check the fuse 	<ol style="list-style-type: none"> 1. Use a power regulator 2. Replace with a new fuse
Some of the air cells are not properly inflated.	<ol style="list-style-type: none"> 1. Is the connection between air cells and the manifold kinked? 2. Is there any air leakage from the air cells? 	<ol style="list-style-type: none"> 1. Check for any kinking between air cells and manifold. 2. Replace new air cell if faulty.

8. EMC Related Notifications

Guidance and manufacturer's declaration – electromagnetic emissions		
The air pump is intended for use in the electromagnetic environment specified below. The customer or the user of the air pump should assure that it is used in such an environment.		
Emissions test	Compliance	Electromagnetic environment – guidance
RF emissions CISPR 11	Group 1	The air pump uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The air pump is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Complies	

Recommended separation distances between portable and mobile RF communications equipment and the air pump			
The air pump is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the air pump can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the air pump as recommended below, according to the maximum output power of the communications equipment.			
Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter m		
	150 kHz to 80 MHz $d = 1,2 \sqrt{P}$	80 MHz to 800 MHz $d = 1,2 \sqrt{P}$	800 MHz to 2,5 GHz $d = 2,3 \sqrt{P}$
0,01	0,12	0,12	0,23
0,1	0,38	0,38	0,73
1	1,2	1,2	2,3
10	3,8	3,8	7,3
100	12	12	23
For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.			
NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.			
NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

Guidance and manufacturer's declaration – electromagnetic immunity


The air pump is intended for use in the electromagnetic environment specified below. The customer or the user of the air pump should assure that it is used in such an environment.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment – guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	±2 kV for power supply lines ±1 kV for input/output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV line(s) to line(s) ±2 kV line(s) to earth	±1 kV line(s) to line(s) ±2 kV line(s) to earth	Mains power quality should be that of a typical commercial or hospital environment.
interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5 % <i>UT</i> (>95 % dip in <i>UT</i>) for 0,5 cycle 40 % <i>UT</i> (60 % dip in <i>UT</i>) for 5 cycles 70 % <i>UT</i> (30 % dip in <i>UT</i>) for 25 cycles <5 % <i>UT</i> (>95 % dip in <i>UT</i>) for 5 sec	<5 % <i>UT</i> (>95 % dip in <i>UT</i>) for 0,5 cycle 40 % <i>UT</i> (60 % dip in <i>UT</i>) for 5 cycles 70 % <i>UT</i> (30 % dip in <i>UT</i>) for 25 cycles <5 % <i>UT</i> (>95 % dip in <i>UT</i>) for 5 sec	Mains power quality should be that of a typical commercial or hospital environment. If the user of the air pump] requires continued operation during power mains interruptions, it is recommended that the air pump be powered from an uninterruptible power supply or a battery.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

NOTE *UT* is the a.c. mains voltage prior to application of the test level.

Guidance and manufacturer's declaration - electromagnetic immunity

The air pump is intended for use in the electromagnetic environment specified below. The customer or the user of the air pump should assure that it is used in such an environment.

<i>Immunity test</i>	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms	Portable and mobile RF communications equipment should be used no closer to any part of the air pump, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance $d = 1,2 \sqrt{P}$ $d = 1,2 \sqrt{P}$ 80 MHz to 800 MHz $d = 2,3 \sqrt{P}$ 800 MHz to 2,5 GHz
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2,5 GHz	3 V/m	where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, ^a should be less than the compliance level in each frequency range. ^b Interference may occur in the vicinity of equipment marked with the following symbol: 

NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures objects and people.

a Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the air pump is used exceeds the applicable RF compliance level above, the air pump should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the air pump.

b Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

9. Warranty

- Prius Healthcare Products guarantees that this equipment is free from defects in material and workmanship. Our obligation under this warranty is limited to the repair of equipment returned to the service address given below, transportation charges prepaid, within 12 months after delivery to the original purchaser for all equipment.
- We agree to service and/or adjust any equipment returned for that purpose and to replace or repair any part, which is proven to be defective at no charge.
- This warranty excludes equipment damage through shipping, tampering, improper maintenance, careless, accident, negligence or misuse, or products which have been altered, repaired or dismantled other than with the manufacture's written authorization and by its approved procedures and by properly qualified technicians.
- In no event shall Prius Healthcare Products be liable for any direct, indirect or consequential damages or losses resulting from the use of equipment.

Prius Healthcare USA

4027 Tampa Road, #3200
Oldsmar, FL 34677, USA

TEL: (813)854-5464
FAX: (813)854-5442